

TECHNICAL INSERT

Item No	585	ChemDisk™ Monitor for	Prepared by	SCW
Edition	8/14/2012	Organic Amines	Approved by	CRM

The ChemDisk™ Personal Monitor is designed to measure exposure to chemicals in order to demonstrate workplace compliance with Permissible Exposure Limits (PELs) and Short Term Exposure Limits (STELs) defined by the Occupational Safety and Health Act of 1970 and Title 29 Code of Federal Regulations.

This **Technical Insert** contains product specific information on use and storage.

9159-585 8/12

Sampling Medium:	Fiberglass coated with 1-naphthyl isothiocyanate (NIT)
Analytes Sampled:	Primary and secondary organic amines; See <i>Sampling Guide</i>
Analytical Method:	Similar to OSHA 60: Desorption in acetonitrile; analysis by high performance liquid chromatography (HPLC) with UV/VIS detection
Recommended Sampling Time:	STEL=15 min; PEL=8 hrs or full shift; Functional range: 15 minutes - 8 hours
Recommended Holding Time:	Monitors must be received by Lab within four (4) weeks after sampling.
Sample Capacity:	Analyte dependent; call for details.
Detection Limit (DL):	Analyte dependent; See <i>Sampling Guide</i>
Sampling Rate (SR):	Typically 7 - 14 mL/min; See <i>Sampling Guide</i>
Interfering Substance(s):	Amines in <i>Sampling Guide</i> list do not interfere if HPLC resolution is adequate.
Effect of Temperature:	Effect on result $\leq 5\%$ within 0 - 50 °C (32 - 122 °F).
Effect of Humidity:	Functions as claimed within 10 - 80% RH.
Accuracy (MTE):	Meets or exceeds OSHA requirements for accuracy: Maximum Total Error (MTE) $\leq 25\%$ at PEL; $\leq 25\%$ at STEL
Storage Conditions:	Store under refrigerated conditions for maximum shelf life. Do Not Use after Expiration Date printed on pouch.
IMPORTANT:	Warranty valid only if Instructions for Use have been followed.
WARNING:	Wash affected area thoroughly if Sampling Medium is contacted.

For detailed directions, see **Instructions for Use** included in each package of Monitors.