
# **SULPHUR DIOXIDE IN FLUE GAS**



## 1. PERFORMANCE

1) Measuring range 0.02-0.3 % Number of pump strokes  $1 (100 \text{m} \, \ell)$ 

2) Sampling time : 1 minute/1 pump stroke

3) Detectable limit : 10 ppm4) Shelf life : 3 years5) Operating temperature  $: 0 \sim 40 \, ^{\circ}\text{C}$ 

6) Reading : Direct reading from the scale calibrated by 1 pump stroke

7) Colour change : White→Orange

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

#### 3. CHEMICAL REACTION

Potassium iodate is reduced.  $SO_2 + KIO_3 + H_2O \rightarrow I_2$ 

## 4. CALIBRATION OF THE TUBE

STANDARD GAS CYLINDER METHOD

### 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Hydrogen sulphide	The accuracy of readings is not affected.	100	Higher readings are given.