
DIETHYL ETHER



1. PERFORMANCE

1) Measuring range 0.04-1.4 % Number of pump strokes $1 (100 \text{m} \ell)$

2) Sampling time : 1.5 minutes/1 pump stroke

3) Detectable limit : 10 ppm4) Shelf life : 3 years5) Operating temperature $: 0 \sim 40 \text{ °C}$

6) Temperature compensation : Necessary (See "TEMPERATURE CORRECTION TABLE") 7) Reading : Direct reading from the scale calibrated by 1 pump stroke

8) Colour change : Orange→Dark green

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

3. CHEMICAL REACTION

Chromium oxide is reduced. $C_2H_5OC_2H_5 + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence	
Acetylene		3%	Whole reagent is changed to Brown.	
Propane		0.2%	"	
Other organic gases and va- pours except Halogenated hydrocarbons	Similar stain is produced.	50	Higher readings are given.	

TEMPERATURE CORRECTION TABLE

Chart	True Concentration (%)					
Readings (%)	0°C (32°F)	10℃ (50°F)	20°C (68°F)	30°C (86°F)	40°C (104°F)	
1.4	1.8	1.6	1.4	1.3	1.1	
1.2	1.5	1.4	1.2	1.1	0.9	
1.0	1.2	1.1	1.0	0.9	0.8	
0.8	0.9	0.9	0.8	0.7	0.6	
0.6	0.6	0.6	0.6	0.6	0.5	
0.5	0.5	0.5	0.5	0.5	0.4	
0.4	0.4	0.4	0.4	0.4	0.3	
0.3	0.3	0.3	0.3	0.3	0.3	