XYLENE



1. PERFORMANCE

1) Measuring range $\begin{array}{c} \text{1.5-1,000 ppm} \\ \text{Number of pump strokes} \end{array}$

2) Sampling time : 4 minutes/2 pump strokes

3) Detectable limit 2 ppm4) Shelf life 1.5 years5) Operating temperature $0 \sim 40 \,^{\circ}\text{C}$

6) Reading : Direct reading from the scale calibrated by 2 pump strokes

7) Colour change : White→Brown

2. RELATIVE STANDARD DEVIATION

RSD-low : 10 % RSD-mid. : 10 % RSD-high : 10 %

3. CHEMICAL REACTION

Iodine pentoxide is reduced. $C_6H_4 (CH_3)_2 + I_2O_5 + H_2SO_4 \rightarrow I_2$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Benzene	Similar stain is produced.	5	Higher readings are given.
Toluene	"	5	"
Methyl alcohol	The accuracy of readings is not affected.	1%	"
Hexane	Pale brown stain is produced.	0.1%	The top of discoloured layer becomes unclear and higher readings are given.