1,2-DICHLOROETHYLENE



1. PERFORMANCE

1) Measuring range : 4.2-84 ppm 9.2-184 ppm 20-400 ppm 42-840 ppm Number of pump strokes 4 (400ml) $2(200 \text{m} \ell)$ $1(100m\ell)$ $1/2(50 \text{m} \ell)$

2) Sampling time : 1.5 minutes/1 pump stroke 3) Detectable limit : 0.5 ppm (4 pump strokes)

4) Shelf life : 1 year (Necessary to store in a refrigerated place : $0 \sim 10^{\circ}$ C)

5) Operating temperature : 0 ~ 40 °C

6) Temperature compensation : Necessary (See "TEMPERATURE CORRECTION TABLE") 7) Reading Direct reading from the scale calibrated by 1 pump stroke : Yellow → Red

8) Colour change

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

3. CHEMICAL REACTION

Hydrogen chloride is produced by an Oxidizer and PH indicator is discoloured.

 $CICH = CHCI + PbO_2 + H_2SO_4 \rightarrow HC1$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Trichloroethylene	Similar stain is produced.	3	Higher readings are given.
Vinyl chloride	"	300	"
Hydrogen chloride	"	10	"
Chlorine	Pale red stain is produced.	15	"

TEMPERATURE CORRECTION TABLE

Scale	True Concentration (ppm)		
Readings (ppm)	5°C (41°F)	10 °C ~ 40 °C (50 ° F ~ 104 ° F)	
400	475	400	
350	415	350	
300	355	300	
250	295	250	
200	235	200	
150	175	150	
100	115	100	
50	55	50	
20	20	20	